

Test Report No. 7191266606-EEC21-WBH
dated 24 Aug 2021



PSB Singapore

**Add value.
Inspire trust.**

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT

Testing of Gloves

CLIENT

Anhui Intco Medical Products Co.,Ltd.
No. 6, Haitang South Road,
Suixi Wuhu Modern Industrial Park,
Suixi County, Huaibei City,
Anhui Province, China

SAMPLE SUBMISSION DATE/ TEST DATE

16 Aug 2021/ 19 Aug 2021

DESCRIPTION OF SAMPLES

S/N	Product Description	Brand/Model	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Disposable Nitrile Exam Gloves (NBR)	Nil	Blue	YL-216388	M	250	Anhui Intco Medical Products Co.,Ltd.

Lot size as specified by client: 100,000 pieces

METHOD OF TEST

As requested by client, the test was conducted as follows:

EN 455-1:2020 Medical gloves for single use
Part 1: Requirements and testing for freedom from hole



Laboratory:
TÜV SÜD PSB Pte. Ltd.
15 International Business Park
TÜV SÜD @ IBP
Singapore 609937

Phone : +65-6778 7777
E-mail: info.sg@tuvsud.com
<https://www.tuvsud.com/sg>
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
15 International Business Park
TÜV SÜD @ IBP
Singapore 609937
TUV[®]

Test Report No. 7191266606-EEC21-WBH
dated 24 Aug 2021




PSB Singapore


RESULTS

Sample: Disposable Nitrile Exam Gloves (NBR), Blue, Lot No. YL-216388, Size M

Clause	Tests	Requirements	AQL	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4 5	Freedom from holes	Shall not leak	1.5	7	200	0	Pass



Yeo Poh Kwang
Associate Engineer



Wong Bee Hui
Product Manager
Medical Health Services (NAM)

APPENDIX



Photo: Disposable Nitrile Exam Gloves (NBR), Blue, Lot No. YL-216388, Size M

Test Report No. 7191266606-EEC21-WBH
dated 24 Aug 2021



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, 15 International Business Park TÜV SÜD @ IBP Singapore 609937.
6. The tests carried out by TÜV SÜD PSB and this report are subject to TÜV SÜD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

Effective 26 January 2021

